

# Initial Guidelines for Manuscripts Employing Data-independent Acquisition Mass Spectrometry for Proteomic Analysis

Robert J. Chalkley<sup>‡§</sup>, Michael J. MacCoss<sup>¶</sup>, Jacob D. Jaffe<sup>||</sup>, and Hannes L. Röst<sup>\*\*</sup>

Proteomic research began largely as an approach for characterizing sample compositions, but most contemporary studies involve a quantitative aspect. Quantification enables comparing sample classes (e.g. healthy *versus* disease) to uncover markers of dysregulation, or comparing protein pull-down experiments to mock pull-downs to determine specific interaction partners. For small-scale comparisons, isotopic labeling, whether introduced metabolically or chemically, is very effective and allows comparison of multiple samples mixed together. However, for comparing a larger number of samples (a dozen or more), label-free strategies are often the most practical option. Reproducible and accurate quantification of a large number of protein and peptide analytes across a large panel of samples remains a singular goal of the proteomics field in general.

Data-independent acquisition mass spectrometry (DIA-MS) is a set of strategies that aim to provide comprehensive coverage and quantification of components in complex peptide mixtures. DIA-MS was developed to circumvent the issues of irreproducible selection of analytes for fragmentation analysis associated with data-dependent acquisition (DDA) and limited analyte coverage (typically  $m/z$  range, most commonly broken down into a series of isolated wide  $m/z$  range windows (1–5)). It has seen considerable growth in the last couple of years as instrumentation that can produce high mass accuracy fragmentation spectra at rates in excess of 10 Hz has become widely available. In parallel to development of acquisition methodologies, new analysis software has also emerged to interpret the resulting data.

*Molecular and Cellular Proteomics* has led the proteomics field in establishing rules for minimum information needed to be provided in submitted manuscripts to evaluate results from different analysis strategies, producing guidelines for authors performing data-dependent MSMS analysis (6), targeted proteomics (7), glycomics/glycoproteomics (8), and clinical proteomic studies (9). These guidelines have in general elevated the standard of published results.

Of late, the journal has published several DIA-MS studies, and it has become evident that even though DIA-MS strate-

gies are still rapidly evolving, a first set of guidelines is required to advise authors on information that should be included in such manuscripts. Hence, in June 2018, the journal organized a meeting of leading researchers in the DIA-MS field in San Diego, CA, to formulate a mutually agreeable set of rules to cover current and anticipated analysis strategies. Representatives from key DIA method, software, and instrument development groups ensured broad community participation. The full list of attendees is provided at the bottom.

The guidelines produced from this meeting were opened to a two-month period of public comment, and the final version is now published (<http://www.mcponline.org/page/DIA-guidelines>) along with this issue of the journal. A companion checklist has also been constructed to assist authors in meeting these guidelines. The journal intends to start implementing these guidelines for relevant manuscripts on March 1.

As DIA-MS methods are still developing, it is anticipated that these guidelines will need to evolve over time to encompass new approaches, but having a first set of guidelines in place will provide a framework for ensuring that results published using these approaches are accountable.

*Acknowledgments*—We would like to thank Steve Carr and Sadiq Zahari for assisting in the organization of the meeting and MCP, Thermo, Waters, and Bruker for providing financial support. The attendees at the meeting were:

Chris Adams, Bruker  
 Nuno Bandeira, UCSD  
 Isabell Bludau, ETH Zürich  
 Andreas Brunner, Max Planck Institute of Biochemistry  
 Al Burlingame, UCSF  
 Steven Carr, Broad Institute (Co-organizer)  
 Robert Chalkley, UCSF (Co-organizer)  
 Meena Choi, Northeastern University  
 Mike Hoopmann, Institute for Systems Biology  
 Jake Jaffe, Broad Institute  
 Brendan MacLean, University of Washington  
 Mike MacCoss, University of Washington  
 Alexey Nesvizhskii, University of Michigan  
 Lukas Reiter, Biognosys  
 Hannes Röst, University of Toronto  
 Birgit Schilling, Buck Institute  
 Brian Searle, Proteome Software

From the <sup>‡</sup>Department of Pharmaceutical Chemistry, University of California San Francisco, San Francisco, California 94143; <sup>¶</sup>Department of Genome Sciences, University of Washington, Seattle, Washington 98195; <sup>||</sup>Proteomics Platform, Broad Institute of MIT and Harvard, Cambridge, Massachusetts 02142; <sup>\*\*</sup>Donnelly Centre, University of Toronto, Toronto, Ontario M5S 3E1, Canada

Stephen Tate, SCIEX  
Stefan Tenzer, Johannes Gutenberg University Mainz  
Hans Vissers, Waters Corporation  
Olga Vitek, Northeastern University  
Juan Antonio Vizcaino, EMBL-EBI  
Sue Weintraub, UT Health San Antonio  
Yue Xuan, Thermo Fischer Scientific  
Saddiq Zahari, ASBMB

§ To whom correspondence should be addressed. E-mail: chalkley@cgl.ucsf.edu.

### REFERENCES

1. Purvine, S., Eppel, J. T., Yi, E. C., and Goodlett, D. R. (2003) Shotgun collision-induced dissociation of peptides using a time of flight mass analyzer. *Proteomics* **3**, 847–850
2. Venable, J. D., Dong, M. Q., Wohlschlegel, J., Dillin, A., and Yates, J. R. (2004) Automated approach for quantitative analysis of complex peptide mixtures from tandem mass spectra. *Nat. Methods* **1**, 39–45
3. Chapman, J. D., Goodlett, D. R., and Masselon, C. D. (2014) Multiplexed and data-independent tandem mass spectrometry for global proteome profiling. *Mass Spectrom. Rev.* **33**, 452–470
4. Egertson, J. D., Kuehn, A., Merrihew, G. E., Bateman, N. W., MacLean, B. X., Ting, Y. S., Canterbury, J. D., Marsh, D. M., Kellmann, M., Zabrouskov, V., Wu, C. C., and MacCoss, M. J. (2013) Multiplexed MS/MS for improved data-independent acquisition. *Nat. Methods* **10**, 744–746
5. Moseley, M. A., Hughes, C. J., Juvvadi, P. R., Soderblom, E. J., Lennon, S., Perkins, S. R., Thompson, J. W., Steinbach, W. J., Geromanos, S. J., Wildgoose, J., Langridge, J. I., Richardson, K., and Vissers, J. P. C. (2018) Scanning quadrupole data-independent acquisition, Part A: Qualitative and quantitative characterization. *J. Proteome Res.* **17**, 770–779
6. Bradshaw, R. A., Burlingame, A. L., Carr, S., and Aebersold, R. (2006) Reporting protein identification data: The next generation of guidelines. *Mol. Cell. Proteomics* **5**, 787–788
7. Abbatiello, S., Ackermann, B. L., Borchers, C., Bradshaw, R. A., Carr, S. A., Chalkley, R., Choi, M., Deutsch, E., Domon, B., Hoofnagle, A. N., Keshishian, H., Kuhn, E., Liebler, D. C., MacCoss, M., MacLean, B., Mani, D. R., Neubert, H., Smith, D., Vitek, O., and Zimmerman, L. (2017) New guidelines for publication of manuscripts describing development and application of targeted mass spectrometry measurements of peptides and proteins. *Mol. Cell. Proteomics* **16**, 327–328
8. Wells, L., and Hart, G. W. (2013) Glycomics: Building upon proteomics to advance glycosciences. *Mol. Cell. Proteomics* **12**, 833–835
9. Celis, J. E., Carr, S. A., and Bradshaw, R. A. (2008) New guidelines for clinical proteomics manuscripts. *Mol. Cell. Proteomics* **7**, 2071–2072