

Supplemental Data

Table 2: Reproducibility of the N-glycosite isolation procedure in conjunction with MRM analysis.

Conc [ng/ml]	Series 1	Series 2	Series 3	Series 4	Mean	% CV
0	0.003	0.002	0.004	0.002	0.003	21%
5	0.008	0.007	0.009	0.008	0.008	7%
7.5	0.012	0.008	0.011	0.012	0.011	17%
50	0.059	0.057	0.071	0.077	0.066	14%
75	0.081	0.080	0.105	0.078	0.086	15%
500	0.447	0.634	0.622	0.872	0.644	27%
750	0.925	0.842	0.889	1.067	0.931	10%

Four serum samples per dilution were subjected to N-glycosite isolation and subsequently analyzed by MRM. Peak height ratios light (protein)/heavy (spiked reference peptide).